Amber® Mill Instructions for Use

HS-IFU-707 Rev.02 (2019. 01. 07)

1. Overview

- · Trade / Device Name : Amber Mill series
- · CommonName: Dental Frame Material for Dental Prosthesis
- Intended Use of the Device: Amber Mill Series are indicated for fabricating glass ceramic restorations such as single-unit anterior and posterior crowns, veneers, inlays/onlays, and anterior 3-unit bridges using CAD/CAM System
- · Classification Name: Porelain Powder for Clinical Use
- · Packaging Unit: Refer to HASS standard package.

2. Instruction



(1) How to Use

This product must be used in accordance with the using methods of a dental CAD/CAM system.

- * Procedure for using glass ceramic blocks *
- Attach the Jig to the accurate location.
- ② Mount it on the CAD/CAM equipment.
- (3) Inputs the size information of the prepared block into the CAD/CAM equipment.
- (4) Inputs correction information needed for processing.
- ⑤ Process the Block using the CAD/CAM equipment.
- 6 Carefully detach the process-completed block from the equipment.
- ⑦ Detach the processed artificial tooth or restoration from the block.
- \$ Artificial teeth or prosthetic separated by heat treatment at 810 \sim 865 \textdegree makes crystallization.
- If necessary, perform stain and glazing treatment.
- (2) Storage and Maintenance before Use
- ① Do not store in package open or dirty place it may contaminate the products.
- 2) Store away from moisture, direct sunlight, and heat.
- 3) Do not reuse or recycle the remaining part once used.

♠ 3. Cautions

- (1) Cautions before Use
 - ① Be careful not to damage the milling tool of the CAD/CAM machine when attaching or detaching the product.
 - 2) Be careful not to get your hand caught in the milling tool.
 - 3 The jig should be attaching to an accurate location.
 - Suppress or remove the dust which may occur during the operation of CAD/CAM machine.
 - ⑤ Do not drop the product on the ground or apply heavy force as it may damage the product.
 - 6 Keep the product out of reach of infants and children.
 - (7) Product should be handled by dental technician.
- (2) Storage and Maintenance before Use
 - Store the product at room temperature in a dry place.
 - 2) Pack and store the product properly to ensure that it is not damaged.

③ Store the product at temperatures ranging from 0° C $\sim 40^{\circ}$ C, in combination with relative humidity of 10° r.H $\sim 90^{\circ}$ r.H, under atmospheric pressures ranging from 500 hPa ~ 1060 hPa.

4. Side effect

It the patient is known to be allergic to any of the components of Amber Mill, the material must not be used to fabricate restorations.

5 Contraindication

- 1) Posterior bridges reaching into the molar region
- ③ Inlay-retained bridges
- (5) Bruxism
- ⑦ Maryland bridges
- Any other use not listed in the indications
- 2) 4-and more-unit bridges
- 4 Very deep sub gingival preparations
- ⑥ Cantilever bridges / extension units

6. Mechanical and Physical Properties

- ① Material: Glass-ceramics
- ② Flexural Strength: over 300 MPa
- 3 Chemical Solubility: below 100 µg/cm²
- 4 Coefficient of Thermal Expansion: 10.0 (±0.5) x 10⁻⁶ K⁻¹
- * This is a single-use product. * Do not reuse.

7. Pictograph



KTR Europe GmbH
Mergenthalerallee 77, 65760 Eschborn, Germany



HASS Corp

77-14, Gwahakdanji-ro, Gangneung-si, Gangwon-do, KOREA 25452 Customer Support: +82-2-2083-1367 E-mail: hasscorp@hassbio.com Website: www.hassbio.com





Amber[®] Mill Recommended Toughening Heat-treatment Schedule

It is possible to differentiate translucency with a single block of Amber[®] Mill. Just decide what shade you will use, then choose the toughening heat-treatment temperature according to your targeted translucency. This will enhance the efficiency in work process and inventory management for CAD/CAM milling blocks.

VITA VACUMAT

Predry ℃	 min.	min.			T °C		min.	VAC min.		
400	3.00	HT	6.50	- 60	HT	815	- 15.00 -	HT	21.50	- 690
		MT	7.05		MT	825		MT	22.05	
		LT	7.20		LT	840		LT	22.20	
		МО	7.40		МО	860		МО	22.40	

^{*} The firing chamber must not be opened during long term cooling.

PROGRAMAT IVOCLAR VIVADENT

B ℃	S min.	t ∕ ℃/ min.	T °C		H min.	VAC. 1 ℃ / VAC. 2 ℃		L °C	tL*
400	3.00	60	HT	815	- 15.00	HT	550/815	690	0
			MT	825		MT	550/825		
			LT	840		LT	550/840		
			MO	860		MO	550/860		

^{*} The firing chamber must not be opened during long term cooling.



